



Ref. Certif. No.

DK-75973-UL

**IEC SYSTEM FOR MUTUAL RECOGNITION OF TEST CERTIFICATES FOR ELECTRICAL EQUIPMENT (IECEE)
CB SCHEME**

CB TEST CERTIFICATE

Product

Conductivity transmitter / Graphical display

Name and address of the applicant

BAUMER A/S
Runetoften 19
Aarhus, 8210 Denmark

Name and address of the manufacturer

BAUMER A/S
Runetoften 19
Aarhus, 8210 Denmark

Name and address of the factory

Note: When more than one factory, please report on page 2

BAUMER A/S
Runetoften 19
Aarhus, 8210 Denmark

[Additional Information on page 2](#)

Ratings and principal characteristics

CombiLyz AFI4 & AFI5: 15-35 Vdc supply.
DFON: 4.5-6.5 Vdc

Trademark (if any)



Type of Customer's Testing Facility (CTF) Stage used

CTF Stage 2

Model / Type Ref.

CombiLyz AFI4 / AFI5, CombiView type DFON

Additional information (if necessary may also be reported on page 2)

National Differences specified in the CB Test Report

[Additional Information on page 2](#)

A sample of the product was tested and found to be in conformity with

IEC 61010-1:2010, IEC 61010-2-201:2013

As shown in the Test Report Ref. No. which forms part of this Certificate

E491206-D1000-1/A0/C0-CB issued on 2018-08-10

This CB Test Certificate is issued by the National Certification Body



-
-
-
-

- UL (US), 333 Pfingsten Rd IL 60062, Northbrook, USA
- UL (Demko), Borupvang 5A DK-2750 Ballerup, DENMARK
- UL (JP), Marunouchi Trust Tower Main Building 6F, 1-8-3 Marunouchi, Chiyoda-ku, Tokyo 100-0005, JAPAN
- UL (CA), 7 Underwriters Road, Toronto, M1R 3B4 Ontario, CANADA

For full legal entity names see www.ul.com/ncbnames

Date: 2018-08-23

Signature:

Jan-Erik Storgaard